



Missouri Department of Health and Senior Services

P.O. Box 570, Jefferson City, MO 65102-0570 Phone: 573-751-6400 FAX: 573-751-6010
RELAY MISSOURI for Hearing and Speech Impaired 1-800-735-2966 VOICE 1-800-735-2466

Gail Vasterling
Director



Jeremiah W. (Jay) Nixon
Governor

Missouri Department of Health and Senior Services Breath Alcohol Program

SIMULATOR CALIBRATION REPORT

This is to certify that the simulator listed below has been examined and tested using standards traceable to the National Institute of Standards and Technology (NIST) in accordance to the standards set by the Rules of Missouri Department of Health and Senior Services, 19 CSR 25-30.

SIMULATOR INFORMATION

Agency: MINER PD
Serial Number: SD2273
Manufacturer: Guth
Model Number: 10-4D

CALIBRATION RESULTS

<u>Reference Temperature</u>	<u>Simulator Temperature</u>
33.99	34.0

This calibration was performed with
NIST-Traceable Thermometer SN: 358440

This simulator was tested by: RWW

This testing was performed: 12/11/2015

This certification expires: 12/11/2016

Signature of certifying DHSS
Scientist:

Name of certifying DHSS Scientist: Ellen Strawsine

MISSOURI SAFETY CENTER SIMULATOR CHECK WORKSHEET



Date: 12/11/15	Time: 1300	Date Last Checked:	Agency	Simulator Serial #:	Simulator reading
Simulator Model:		<input type="checkbox"/> 2100 <input checked="" type="checkbox"/> 10-4D <input type="checkbox"/> 12V500	Email address:	SD 2273	
Thermometer serial #:		358440	Agency property #:	LASH @ CITY OF MINER . com	
Thermometer certification date:		09/08/15	Thermometer reading		
Thermometer expiration date:		09/08/16			
1 st check time:		1325			
2 nd check time:		1327			
3 rd check time:		1329			
4 th check time:		1331			
5 th check time:		1333			
Average readings:					
Bias calculation:		- .01			

This form meets or exceeds the requirements of the Missouri Department of Health breath alcohol program.

☒ Check "o" rings on quick-disconnects and replace as needed.

☒ Check simulator "o" ring and replace as needed.

☒ Check jar for breaks/cracks and replace as needed.

COMMENTS:



SIMULATOR SERIAL NO.: SD2273
 EXPIRATION DATE: 12/01/2016
 DATE OF CALIBRATION CHECK: 12/11/2015
 NIST REF. THEM. SERIAL NO.: 0358440
 AVERAGE SIM. TEMP: 33.99 C
 ANALYST INITIALS: RWW

This simulator is operating within DHSS breath alcohol specifications (PROC 23-20.911).

TECHNICIAN INITIALS: RWW